

Package ‘CLSI EP15’

January 20, 2025

Type Package

Title Clinical and Laboratory Standards Institute (CLSI) EP15-A3
Calculations

Version 0.1.0

Maintainer Claudio Antonio Rank Filho <claudio.filho@hitechologies.com.br>

Description Calculations of ``EP15-A3 document. A manual for user verification of precision and estimation of bias'' CLSI (2014, ISBN:1-56238-966-1).

License MIT + file LICENSE

Encoding UTF-8

LazyData true

RoxygenNote 7.2.3

Depends R (>= 4.0)

Imports stats, dplyr, tidyr

VignetteBuilder knitr

Suggests knitr, rmarkdown

NeedsCompilation no

Author Claudio Antonio Rank Filho [aut, cre]

Repository CRAN

Date/Publication 2023-11-10 19:43:23 UTC

Contents

bias_validation_interval	2
calculate_aov_infos	3
calculate_bias_interval	3
calculate_dfWL	4
calculate_df_combined	5
calculate_F_uvl	6
calculate_m	6
calculate_n0	7

calculate_se_c	7
calculate_se_rm	8
calculate_se_rm_a_lowerupper	8
calculate_se_rm_a_u	9
calculate_se_rm_a_Ucoverage	9
calculate_se_rm_a_Uk	10
calculate_se_rm_scenario_b_c	10
calculate_se_rm_scenario_d_e	11
calculate_se_x	11
calculate_uvl_info	12
create_table_ep_15	12
dfc_references	13
ferritin_long	13
ferritin_wider	14

Index	15
--------------	-----------

bias_validation_interval
Calculate bias validation interval

Description

Calculate bias validation interval

Usage

```
bias_validation_interval(TV, m, se_c)
```

Arguments

TV	True value
m	factor
se_c	SE Combined

Value

named list with the interval

calculate_aov_infos *Calculate ANOVA Results and Imprecision Estimates*

Description

Calculate ANOVA Results and Imprecision Estimates

Usage

```
calculate_aov_infos(ep_15_table)
```

Arguments

ep_15_table table generated from create_table_ep_15()

Value

Named list with ANOVA Results and Imprecision Estimates

Examples

```
calculate_aov_infos(create_table_ep_15(CLSIEP15::ferritin_long, data_type = 'long'))
```

calculate_bias_interval

Calculate bias interval from TV

Description

Calculate bias interval from TV

Usage

```
calculate_bias_interval(  
  scenario,  
  nrun,  
  nrep,  
  SWL,  
  SR,  
  nsamples,  
  expected_mean,  
  user_mean,  
  ...  
)
```

Arguments

scenario	Choosed scenario from section 3.3 of EP15-A3
nrun	Number of runs
nrep	number of repetitions per run (n_0)
SWL	S within laboratory (obtained from anova)
SR	S repeatability (obtained from anova)
nsamples	total number of samples tested usual 1
expected_mean	Expected mean or TV
user_mean	Mean of all samples (obtained from anova)
...	additional parameters necessary for processing the choosed scenario

Value

a named list with the defined mean, the interval significance (user mean should be in for approval), and total bias (user mean - TV)

Examples

```
calculate_bias_interval(scenario = 'E',
nrun = 7,
nrep = 5,
SWL = .042,
SR = .032,
nsamples = 2,
expected_mean = 1,
user_mean = .94
)
```

calculate_dfWL

Calculate degrees of freedom within-lab as specified in appendix B

Description

Calculate degrees of freedom within-lab as specified in appendix B

Usage

```
calculate_dfWL(cvr_manufacture, cvwl_manufacture, k, n0, N)
```

Arguments

cvr_manufacture	CV repeatability informed by the manufacturer
cwvl_manufacture	CV within-lab informed by the manufacturer
k	the number of runs
n0	the “average” number of results per run
N	the total number of replicates

Value

dfwl

calculate_df_combined *Calculate degrees of freedom of SE C (SE combined) given a selected scenario and additional parameters necessary for the scenario*

Description

Calculate degrees of freedom of SE C (SE combined) given a selected scenario and additional parameters necessary for the scenario

Usage

calculate_df_combined(scenario, ...)

Arguments

scenario	Scenario (A, B, C, D, E)
...	additional parameters necessary for the scenario

Value

DF

`calculate_F_uv1` *Calculate the UVL factor*

Description

Calculate the UVL factor

Usage

```
calculate_F_uv1(nsamp = 1, df, alpha = 0.05)
```

Arguments

<code>nsamp</code>	n samples in the study
<code>df</code>	degrees of freedom
<code>alpha</code>	confidence level

Value

Uvl factor

`calculate_m` *Calculate M*

Description

Calculate M

Usage

```
calculate_m(df, conf.level = 95, nsamples = 1)
```

Arguments

<code>df</code>	degrees of freedom
<code>conf.level</code>	confidence interval
<code>nsamples</code>	number of samples

Value

m factor

<code>calculate_n0</code>	<i>Calculate n0</i>
---------------------------	---------------------

Description

Calculate n0

Usage

```
calculate_n0(long_result_table)
```

Arguments

<code>long_result_table</code>	table generated by <code>create_table_ep_15</code> function
--------------------------------	---

Value

The n0 number which refers to Number of Results per Run

<code>calculate_se_c</code>	<i>Calculate SE combined based on SE X and SE RM</i>
-----------------------------	--

Description

Calculate SE combined based on SE X and SE RM

Usage

```
calculate_se_c(se_x, se_rm)
```

Arguments

<code>se_x</code>	SE X
<code>se_rm</code>	SE RM

Value

SE C

calculate_se_rm

Calculate SE RM given a scenario and a list of additional args that can change based on the selected scenario or sub scenario

Description

Calculate SE RM given a scenario and a list of additional args that can change based on the selected scenario or sub scenario

Usage

```
calculate_se_rm(scenario, additional_args)
```

Arguments

scenario	scenario (A, B, C, D, E)
additional_args	additional arguments list

Value

SE RM

calculate_se_rm_a_lowerupper

Calculate SE RM for scenario A when f the manufacturer supplies lower and upper limits and coverage confidence interval (95 or 99...)

Description

Calculate SE RM for scenario A when f the manufacturer supplies lower and upper limits and coverage confidence interval (95 or 99...)

Usage

```
calculate_se_rm_a_lowerupper(upper, lower, coverage)
```

Arguments

upper	upper limit
lower	lower limit
coverage	coverage

Value

SE RM

`calculate_se_rm_a_u`

Calculate SE RM for scenario A when “standard error” or “standard uncertainty” (abbreviated by lowercase “u”) or “combined standard uncertainty” (often denoted by “uC ”)

Description

Calculate SE RM for scenario A when “standard error” or “standard uncertainty” (abbreviated by lowercase “u”) or “combined standard uncertainty” (often denoted by “uC ”)

Usage

```
calculate_se_rm_a_u(u)
```

Arguments

u	“standard error” or “standard uncertainty” (abbreviated by lowercase “u”) or “combined standard uncertainty” (often denoted by “uC ”)
---	---

Value

SE RM

`calculate_se_rm_a_Ucoverage`

Calculate SE RM for scenario A when f the manufacturer supplies an “expanded uncertainty” (abbreviated by uppercase “U”) for the TV and coverage e.g. 95 or 99,

Description

Calculate SE RM for scenario A when f the manufacturer supplies an “expanded uncertainty” (abbreviated by uppercase “U”) for the TV and coverage e.g. 95 or 99,

Usage

```
calculate_se_rm_a_Ucoverage(U, coverage)
```

Arguments

U	expanded uncertainty
coverage	coverage

Value

SE RM

`calculate_se_rm_a_Uk` *Calculate SE RM for scenario A when f the manufacturer supplies an “expanded uncertainty” (abbreviated by uppercase “U”) for the TV and the “coverage factor” (abbreviated by “k”)*

Description

Calculate SE RM for scenario A when f the manufacturer supplies an “expanded uncertainty” (abbreviated by uppercase “U”) for the TV and the “coverage factor” (abbreviated by “k”)

Usage

```
calculate_se_rm_a_Uk(U, k)
```

Arguments

U	expanded uncertainty
k	coverage factor

Value

SE RM

`calculate_se_rm_scenario_b_c`

Calculate SE RM for scenario B or C If the reference material has a TV determined by PT or peer group results

Description

Calculate SE RM for scenario B or C If the reference material has a TV determined by PT or peer group results

Usage

```
calculate_se_rm_scenario_b_c(sd_rm, nlab)
```

Arguments

sd_rm	SD RM
nlab	number of lab or peer group results

Value

SE RM

`calculate_se_rm_scenario_d_e`

Calculate SE RM for scenario D or E If the TV represents a conventional quantity value or When working with a commercial QC material supplied with a TV for which the standard error cannot be estimated

Description

Calculate SE RM for scenario D or E If the TV represents a conventional quantity value or When working with a commercial QC material supplied with a TV for which the standard error cannot be estimated

Usage`calculate_se_rm_scenario_d_e()`**Value**

SE RM

`calculate_se_x` *Calculate SE x*

Description

Calculate SE x

Usage`calculate_se_x(nrun, nrep, SWL, SR)`**Arguments**

nrun	Run number
nrep	Number of repetitions per run n0
SWL	SWL from aov table
SR	SR from aov table

Value

SE X

`calculate_uvl_info` *Calculate upper verification limit*

Description

Generic function for calculating UVL the return is a named list and cv_uvl_r and cv_uvl_wl depends on what is the input (S or CV) if the input is SR and SWL the returns is S

Usage

```
calculate_uvl_info(aov_return, nsamp = 1, cvr_or_sr, cvwl_or_sw1)
```

Arguments

<code>aov_return</code>	Return of calculate_aov_info()
<code>nsamp</code>	number of samples in the experiment
<code>cvr_or_sr</code>	Desirable CV or S repeatability
<code>cvwl_or_sw1</code>	Desirable CV or S within-lab

Value

Named list with UVL params

Examples

```
data <- create_table_ep_15(ferritin_wider)
aov_t <- calculate_aov_infos(data)
calculate_uvl_info(aov_t, nsamp = 5, cvr_or_sr = .43, cvwl_or_sw1 = .7)
```

`create_table_ep_15` *Create table for precision calculations*

Description

Create table for precision calculations

Usage

```
create_table_ep_15(data, data_type = "wider")
```

Arguments

<code>data</code>	a long or a wider data.frame with the same structure of CLSIEP15::ferritin_long or CLSIEP15::ferritin_wider
<code>data_type</code>	c('wider', 'long')

Value

a data.frame with renamed columns and structure adjustments

Examples

```
data <- create_table_ep_15(ferritin_long, data_type = "longer")
```

dfc_references

Reference of degrees of freedom based on tau given in the CLSI Manual

Description

Reference of degrees of freedom based on tau given in the CLSI Manual

Usage

```
dfc_references
```

Format

‘dfc_references’ A data frame with 390 rows and 4 columns:

tau tau

df degrees of freedom

labs number of labs or peers

runs number of runs ...

Source

CLSI EP15-A3

ferritin_long

Ferritin data used in CLSI document examples in wide format

Description

Ferritin data used in CLSI document examples in wide format

Usage

```
ferritin_long
```

Format

‘ferritin_long‘ A data frame with 25 rows and 3 columns:

rep Repetition of sample

name Run of the Runs obtained from 5 distinct days

value result of the observation ...

Source

CLSI EP15-A3

ferritin_wider

Ferritin data used in CLSI document examples in wide format

Description

Ferrin data used in CLSI document examples in wide format

Usage

ferritin_wider

Format

‘ferritin_wider‘ A data frame with 5 rows and 6 columns:

rep Repetition of sample

Run_1, Run_2, Run_3, Run_4, Run_5 Runs from 5 distinct days ...

Source

CLSI EP15-A3

Index

* datasets

dfc_references, 13

ferritin_long, 13

ferritin_wider, 14

bias_validation_interval, 2

calculate_aov_infos, 3

calculate_bias_interval, 3

calculate_df_combined, 5

calculate_dfWL, 4

calculate_F_uv1, 6

calculate_m, 6

calculate_n0, 7

calculate_se_c, 7

calculate_se_rm, 8

calculate_se_rm_a_lowerupper, 8

calculate_se_rm_a_u, 9

calculate_se_rm_a_Ucoverage, 9

calculate_se_rm_a_Uk, 10

calculate_se_rm_scenario_b_c, 10

calculate_se_rm_scenario_d_e, 11

calculate_se_x, 11

calculate_uv1_info, 12

create_table_ep_15, 12

dfc_references, 13

ferritin_long, 13

ferritin_wider, 14